

# Single-shot detection of bacterial endospores via coherent Raman spectroscopy

Dmitry Pestov<sup>†‡</sup>, Xi Wang<sup>†</sup>, Gombojav O. Ariunbold<sup>†</sup>, Robert K. Murawski<sup>†§</sup>, Vladimir A. Sautenkov<sup>†</sup>, Arthur Dogariu<sup>¶</sup>, Alexei V. Sokolov<sup>†</sup>, and Marlan O. Scully<sup>†¶||</sup>

<sup>†</sup>Institute for Quantum Studies and Departments of Physics and Chemical Engineering, Texas A&M University, College Station, TX 77843; and <sup>¶</sup>Applied Physics and Materials Science Group, Engineering Quadrangle, Princeton University, Princeton, NJ 08544

Contributed by Marlan O. Scully, November 13, 2007 (sent for review September 17, 2007)

**Recent advances in coherent Raman spectroscopy hold exciting promise for many potential applications. For example, a technique, mitigating the nonresonant four-wave-mixing noise while maximizing the Raman-resonant signal, has been developed and applied to the problem of real-time detection of bacterial endospores. After a brief review of the technique essentials, we show how extensions of our earlier experimental work [Pestov D, *et al.* (2007) *Science* 316:265–268] yield single-shot identification of a small sample of *Bacillus subtilis* endospores ( $\approx 10^4$  spores). The results convey the utility of the technique and its potential for “on-the-fly” detection of biohazards, such as *Bacillus anthracis*. The application of optimized coherent anti-Stokes Raman scattering scheme to problems requiring chemical specificity and short signal acquisition times is demonstrated.**

biohazard | CARS | hybrid | technique | identification

In recent years, local as well as remote real-time detection and identification of warfare agents has become vital. Whereas the detection of warfare chemical compounds, such as nerve gases (1) and explosives (2), has been developed, there is still an apparent need for satisfactory detection technology targeting biohazards, like *Bacillus anthracis*. A number of optical techniques have been considered as alternatives to well established but cumbersome and time-consuming biological methods (for a review, see ref. 3). In particular, the Raman spectra of endospores have been characterized as their “fingerprints.” Detection schemes based on conventional spontaneous Raman spectroscopy (4–9) and surface-enhanced Raman spectroscopy (SERS) (see, e.g., ref. 10) have been reported. For  $1.4 \times 10^3$  spores, a data collection time as low as 10 s has been achieved (11). In this article, we consider a complementary approach, coherent anti-Stokes Raman scattering (CARS) spectroscopy (12, 13), and demonstrate the acquisition of a CARS spectrum from  $\approx 10^4$  *Bacillus subtilis* spores in a single laser shot.

The use of CARS for spore detection was suggested several years ago (14). A set of femtosecond adaptive spectroscopic techniques for CARS (FAST CARS), including electronic-resonance-enhanced coherent Raman scattering (15, 16), laser pulse shaping (17), and timing, was proposed and analyzed. The authors showed how to boost the CARS efficiency and discussed possible strategies for real-time bacterial spore detection.

Here, we use a recently introduced time-frequency adaptation of CARS that brings together the concepts of frequency-resolved (18–22) and time-resolved (23–25) CARS measurements. The scheme combines broadband Raman excitation by a pair of ultrashort laser pulses and a time-delayed narrowband probing of the resultant coherent molecular vibrations (Fig. 1*a*). The broadband preparation allows the monitoring of multiple Raman lines in a single measurement—i.e., gives specificity. The narrowband probing allows for frequency-resolved acquisition. Recording of the whole CARS spectrum at once, without tuning the laser wavelengths, makes the technique relatively insensitive to fluctuations. As we explain in the following section, the technique provides a means to differentiate between the reso-

nant and nonresonant (NR) contributions. It also helps to mitigate the strength of NR four-wave mixing (FWM). The use of a non-zero probe delay further optimizes the ratio between the two contributions in favor of the vibrationally resonant signal, which we will simply refer to as CARS.

We have recently used this broadband preparation—narrowband time-delayed probing scheme (called hybrid CARS for short)—to obtain the spectrum from highly scattering samples of sodium dipicolinate powder and *Bacillus subtilis* endospores (26). The same approach was applied by Stauffer *et al.* (27) to CARS on neat solutions of toluene and chloroform as well as rhodamine 6G with an emphasis on prospective monitoring of intramolecular vibrational energy redistribution; another but conceptually similar scheme was also implemented by Hamaguchi *et al.* (28). By delaying a narrowband pump pulse relative to dispersion-compensated supercontinuum (used as a Stokes pulse within a standard broadband CARS scheme), they observed NR background-free CARS signal from impulsively excited vibrational modes of indene. Finally, the time-resolved dynamics of the resonant and NR broadband picosecond CARS signals from gas-phase nitrogen was investigated in ref. 29.

In this contribution, we improve on our hybrid CARS measurements on *B. subtilis* endospores (26) and elaborate on the optimization of the initially proposed experimental setup. First, the wavelengths of the pump, Stokes, and probe beams are shifted into the near-IR domain, where the photodamage threshold for the spores is higher as compared with the visible-range wavelengths. Second, the bandwidth and delay of the probe pulse are carefully optimized. These steps have led to a  $10^2$  to  $10^3$  increase in the sensitivity of our CARS setup and enabled us to acquire the spectrum from a small sample volume ( $\approx 10^4$  spores) in a single laser shot.

Because CARS efficiency is expected to be proportional to the intensity of the excitation laser pulses, the generated signal can in principle be increased (and the required sample size decreased) by tighter focusing of the applied laser beams. However, there is obviously a limit imposed by the laser-induced damage to the sample. In our previous work, wherein signal collection was done over many laser shots, the applied laser intensities (and therefore the average signal strengths) were limited by cumulative damage to the sample. An obvious question that one may ask

Author contributions: D.P., A.V.S., and M.O.S. designed research; D.P., X.W., G.O.A., and A.D. performed research; R.K.M., V.A.S., and M.O.S. contributed new reagents/analytic tools; D.P., X.W., G.O.A., R.K.M., V.A.S., A.D., A.V.S., and M.O.S. analyzed data; and D.P., A.V.S., and M.O.S. wrote the paper.

The authors declare no conflict of interest.

Freely available online through the PNAS open access option.

<sup>†</sup>To whom correspondence may be addressed at: Department of Physics, MS 4242, Texas A&M University, College Station, TX 77843-4242. E-mail: [dmip@neo.tamu.edu](mailto:dmip@neo.tamu.edu).


<sup>§</sup>Present address: Department of Physics, Drew University, Madison, NJ 07940.

<sup>||</sup>To whom correspondence may be addressed. E-mail: [scully@tamu.edu](mailto:scully@tamu.edu).

This article contains supporting information online at [www.pnas.org/cgi/content/full/0710427105/DC1](http://www.pnas.org/cgi/content/full/0710427105/DC1).

© 2008 by The National Academy of Sciences of the USA





**Fig. 2.** The calculated spectral bandwidth (a) and peak spectral intensity (b) of the resonant (CARS) and nonresonant (NR FWM) contributions into  $|P^{(3)}|^2$ , without taking into account the interference term,  $2 \operatorname{Re}(P_{\text{NR}}^{(3)} P_{\text{R}}^{(3)})$ . The parameters used are  $\chi_{\text{NR}}^{(3)} = 1$ ,  $\alpha = 1$ ,  $\gamma = 1$ ,  $\Delta\omega_1 = \Delta\omega_2 = 50\gamma$ ,  $\tau = 0$ ,  $S_0 = 1$ , and  $A_3 = 1$ .


contributions. Furthermore, cutting the probe spectrum facilitates substantial reduction of the NR background without appreciable loss in the peak spectral intensity of the CARS signal (see Fig. 2b). Please note that the relative position of the two curves along the vertical axis in Fig. 2b depends on the choice of  $\chi_{\text{NR}}^{(3)}$ , the Raman scattering coefficient  $\alpha$  (see Eq. 2), and the molecular vibration decay rate  $\gamma$ . In particular, they happen to merge in the limit  $\Delta\omega_3 \ll \gamma$  because  $\gamma\chi_{\text{NR}}^{(3)}/\alpha = 1$  for the demonstrated calculations.

One can obtain the following analytical expression for the ratio of the two peaks at  $\omega = \omega_{\text{c3}} + \Omega_{\text{R}}$ :

$$R = \left| \frac{P_{\text{R}}^{(3)}(\omega_{\text{c3}} + \Omega_{\text{R}})}{P_{\text{NR}}^{(3)}(\omega_{\text{c3}} + \Omega_{\text{R}})} \right|^2 = \frac{2\pi \ln(2)}{W^2} \left| \frac{\alpha}{\chi_{\text{NR}}^{(3)}} \right|^2 \exp\left(4 \ln(2) \frac{\gamma^2}{W^2}\right) \cdot \left(1 - \operatorname{erf}\left(\sqrt{2 \ln(2)} \frac{\gamma}{W}\right)\right)^2. \quad [6]$$

Here,  $W \equiv \Delta\omega_1 \Delta\omega_3 / (\Delta\omega_1^2 + \Delta\omega_3^2)^{1/2}$  and  $\operatorname{erf}(x) \equiv (2/\pi^{1/2}) \int_0^x \exp(-t^2) dt$  is the error function. In the limit of a large probe bandwidth [i.e., for  $\Delta\omega_3 \gg \gamma$  (we also assume  $\Delta\omega_1 \gg \gamma$ )],  $R \approx 2\pi \ln(2)/W^2 |\alpha/\chi_{\text{NR}}^{(3)}|^2$ . In the opposite limit (i.e., for  $\Delta\omega_3 \ll \gamma$ ), one gets  $R \approx 1/\gamma^2 |\alpha/\chi_{\text{NR}}^{(3)}|^2$ . Finally,  $R$  scales as  $\Delta\omega_3^{-2}$  for  $\gamma \ll \Delta\omega_3 \ll \Delta\omega_1$ .

The other parameter that has been left out of the discussion so far, but that can be used to further improve the signal-to-background ratio, is the probe pulse delay. For any fixed probe pulse bandwidth, the NR contribution peaks at zero probe delay and decreases as the overlap between the three pulses is reduced via the probe pulse timing (see Fig. 3a). The resonant response, in contrast, becomes maximal at some positive probe delay and decreases at larger delays with a decay rate determined by the dephasing time of molecular oscillations [Fig. 3b; see also supporting information (SI) Fig. 9]. When normalized



**Fig. 3.** The calculated magnitude of  $|P_{\text{NR}}^{(3)}|^2$  (a) and  $|P_{\text{R}}^{(3)}|^2$  (b) at  $\omega = \omega_{\text{c3}} + \Omega_{\text{R}}$  as a function of the probe spectral bandwidth  $\Delta\omega_3$  and time delay  $\tau$  relative to the overlapped preparation pulses. The parameters are similar to those used for Fig. 2, but  $\tau$  is varied and the probe pulse amplitude is normalized such that  $A_3 = (4 \ln(2)/\pi)^{1/4} \Delta\omega_3^{-1/2}$  to have the probe pulse energy constant.

on the probe pulse energy, the resonant contribution reaches maximum for the probe pulse bandwidth  $\Delta\omega_3$  on the order of  $2\gamma$ , and its delay  $\tau$  about  $\Delta\omega_3^{-1}$ . The NR FWM peaks at zero probe delay and  $\Delta\omega_3 \sim \Delta\omega_1$  (which is out of the range plotted in Fig. 3a).


Simulations presented in Fig. 4 summarize the effect of the probe bandwidth and delay on the spectrally resolved profiles of  $|P_{\text{NR}}^{(3)}|^2$ ,  $|P_{\text{R}}^{(3)}|^2$ , and  $|P_{\text{NR}}^{(3)} + P_{\text{R}}^{(3)}|^2$ . The initially featureless spectrum, dominated by the NR FWM (Fig. 4a–c), reveals the presence of a Raman-resonant response, distinct from the interfering NR background, when the probe pulse bandwidth is adjusted (see Fig. 4d–f); adding probe delay further improves the signal-to-background ratio, as shown in Fig. 4g–i. The actual optimal values of the probe pulse parameters depend on  $\gamma$ , the ratio  $\gamma\chi_{\text{NR}}^{(3)}/\alpha$ , the noise level, and the necessary spectral resolution.

It is worthwhile to note that suitable probe pulse shapes are not limited to the considered Gaussian waveform. For example, one can use a probe pulse with a  $\operatorname{sinc}^2(\Delta\omega_3 t/2)$ -shaped temporal profile, as we do in our experiment. By putting the preparation pulses on the node of the probe field, as shown in SI Fig. 10, one can virtually cancel out the NR FWM contribution at a finite probe pulse delay. Of course, the timing between the pulses is essential here, and multiple scattering inside the sample, if present, fills in the field node to some extent.

## Experimental Results and Discussion

An essential improvement in performance was achieved by shifting of the pump, Stokes, and probe wavelengths from the visible into the IR domain. This allowed us to increase the energy of the pulses and their peak intensities by three to five times without destroying the spores.






**Fig. 4.** The calculated spectrally resolved profiles of  $|P_{RI}^{(3)}|^2$  (Left),  $|P_{RI}^{(3)}|^2$  (Center), and  $|P_{RI}^{(3)} + P_{RI}^{(3)}|^2$  (Right) for  $\Delta\omega_3 = 50\gamma$ ,  $\tau = 0$  (a–c);  $\Delta\omega_3 = 2\gamma$ ,  $\tau = 0$  (d–f); and  $\Delta\omega_3 = 2\gamma$ ,  $\tau = 2/\gamma$  (g–i). The other parameters are the same as in Fig. 2.

The next step toward superior sensitivity was the optimization of the spectral bandwidth and the time delay of the probe pulse. Because the observed NR background in the recorded CARS spectra was dominated by the multiple scattering within the sample, this was done experimentally. A set of spectrograms, shown in Fig. 5 Left, was recorded for different spectral bandwidths of the probe pulse. For each probe delay in a spectrogram, the acquired spectrum was processed to remove the NR background and account for the pump–Stokes excitation profile.<sup>††</sup> The normalized spectrograms (see Fig. 5 Right) reveal sharp resonant features on top of the smooth NR background as well as the optimal probe delay  $\tau_{\text{opt}}$ , corresponding to the maximal ratio between CARS and NR FWM, for every chosen value of  $\Delta\omega_3$ . Fig. 6 illustrates the recorded, flat-field corrected CARS spectra at  $\tau = 0$  and  $\tau = \tau_{\text{opt}}$ .<sup>‡‡</sup> As expected, two tendencies can be observed. If the probe bandwidth is small compared with the Raman line width, the visibility of the Raman peak depends only weakly on the probe delay, and the signal magnitude is low. If the probe bandwidth is large, the signal is stronger but the contrast degrades. For the observed Raman transitions,  $\Delta\omega_3 \approx 30 \text{ cm}^{-1}$  and the time delay  $\tau \approx 0.77 \text{ ps}$  were found to be optimal (see Figs. 5d and 6c).

Once we had the parameters optimized, the setup performance was evaluated by decreasing the acquisition time. Fig. 7 shows CARS spectra of *B. subtilis* spores taken in only one to three laser shots. Spore Raman transitions in the fingerprint region ( $1,300\text{--}1,700 \text{ cm}^{-1}$ ) are apparent even after a single shot. For these measurements, the laser system was switched from normal (continuous) operation at 1 kHz rep rate into the single-shot mode, and the laser pulses were fired manually. The integration time of the charge-coupled device (CCD) camera


was set to 3 s to ensure a comfortable time span for the manual firing of up to 10 shots. The variation of the CCD integration time from 0.5 to 10 s did not show any appreciable change in the noise level of the single-shot spectra because the dominant part of the noise came from the scattered laser radiation entering the



**Fig. 5.** The experimentally obtained CARS spectrograms (Left) and processed CARS-vs.-FWM profiles (Right) acquired with a pellet of *B. subtilis* spores for different FWHM spectral bandwidths of the probe pulse  $\Delta\omega_3$ :  $40 \text{ cm}^{-1}$  (a and b),  $30 \text{ cm}^{-1}$  (c and d),  $20 \text{ cm}^{-1}$  (e and f),  $15 \text{ cm}^{-1}$  (g and h), and  $10 \text{ cm}^{-1}$  (i and j). The integration time is 1 s. The central wavelengths of the pump, Stokes, and probe pulses are  $1.28 \mu\text{m}$ ,  $1.62 \mu\text{m}$ , and  $805 \text{ nm}$ , respectively. The energy of the probe pulse is kept constant while the spectral bandwidth is adjusted.

<sup>††</sup>We have found it useful to obtain a smoothed replica of the acquired CARS spectrum by averaging the recorded counts over 200 adjacent pixels and then dividing the original spectrum by the generated profile. It is analogous to forming the ratio of CARS plus NR FWM over NR FWM.

<sup>‡‡</sup>The etaloning effect, common for back-illuminated CCDs in the spectral range of interest, leads to inhomogeneous optical response of the CCD over its surface. It results in parasitic amplitude modulation of recorded spectra. To correct for ghost spectral peaks, we acquired a sample spectrum with the probe beam blocked (high harmonic generation due to the pump and Stokes fields contributed as a smooth background) and then divided all CARS spectra by the obtained profile.




**Fig. 6.** Cross-sections of the recorded CARS spectrograms, similar to those in Fig. 5, for two different probe delays: zero delay (Left) and optimal delay (Right). The last one is adjusted for different values of  $\Delta\omega_3$ : 60  $\text{cm}^{-1}$  (a), 40  $\text{cm}^{-1}$  (b), 30  $\text{cm}^{-1}$  (c), 20  $\text{cm}^{-1}$  (d), 15  $\text{cm}^{-1}$  (e), and 10  $\text{cm}^{-1}$  (f).

spectrometer. At last, the reported pulse energies were deduced from the power measurements in the normal (continuous) operation mode. We found that they differed from the actual pulse energies when the laser system was operated in the single-shot mode. The ratio of the single-shot to the continuous-mode pulse energy was 0.75, 0.47, and 1.3 for the pump, Stokes, and probe pulses, respectively.


Note that the single-shot spectra were obtained from a fairly small number of spores. Taking the effective laser pulse penetration depth for the highly scattering sample (a clump of closely packed spores) to be 10  $\mu\text{m}$  and the diameter of the probed region to be 32  $\mu\text{m}$ , one gets a rough estimate for the probed volume of  $10^4 \mu\text{m}^3$ . Allocating 1  $\mu\text{m}^3$  for a single spore gives  $\approx 10^4$  contributing spores.

The dependence of the resonant CARS signal on the energy of the pump and Stokes pulses is shown in Fig. 8. Here, the resonant contribution is defined as an integral over the two Raman peaks, 1,395  $\text{cm}^{-1}$  and 1,445  $\text{cm}^{-1}$ . The data reveal that at high enough pulse energies, the dependence strongly deviates from the expected linear increase. The CARS signal magnitude saturates and even drops down at higher pulse energies. The total acquired signal, which is dominated by the nonresonant contribution, exhibits similar dependence on the pulse energies but with a slightly weaker decay at the high-energy limit (data not shown). Within the available range (up to 3  $\mu\text{J}$  per pulse), we observed a linear dependence of the CARS signal on the energy of the probe pulse.

The saturation and decrease of the generated CARS signal as a function of the pump or Stokes pulse energies is probably due to the laser-induced damage of the exposed endospores.



**Fig. 7.** Acquired CARS spectra for one, two, and three laser shots at the optimal probe delay. Parameters were as follows. (a) Pump wavelength  $\lambda_1 = 1.25 \mu\text{m}$ , 4  $\mu\text{J}$  per pulse; probe,  $\lambda_3 = 805.8 \text{ nm}$ ,  $\Delta\omega_3 = 30 \text{ cm}^{-1}$ , 3  $\mu\text{J}$  per pulse; Stokes,  $\lambda_2 = 1.54 \mu\text{m}$ , 4  $\mu\text{J}$  per pulse. (b) Stokes,  $\lambda_2 = 1.56 \mu\text{m}$ , 4  $\mu\text{J}$  per pulse.



**Fig. 8.** Dependence of the CARS signal on the energy of the pump (a) and Stokes (b) pulses. Other parameters are  $\lambda_1 = 1.24 \mu\text{m}$ ,  $\lambda_2 = 1.54 \mu\text{m}$ ,  $\lambda_3 = 805.7 \text{ nm}$ , and  $\Delta\omega_3 = 30 \text{ cm}^{-1}$ . The CARS signal was recorded in the single-shot regime, moving the sample after each laser shot to avoid accumulated effects due to the laser-induced damage.

Accounting for the beam focusing and the scaling factors, we estimate the damage threshold for our spore sample to be 0.2  $\text{J}/\text{cm}^2$  per ultrashort pump or Stokes pulse. For 50-fs pulses, this energy fluence corresponds to a peak intensity of  $\approx 3 \times 10^{12} \text{ W}/\text{cm}^2$ . Note that this threshold is retrieved from the single-shot measurements and corresponds to ultrafast (femtosecond-scale) photodamage. We have also observed a longer-scale cumulative effect of CARS signal degradation as a function of the shot number. To avoid the substantial damage of the sample over 100 laser shots, one needs to use a pulse energy fluence that is at least three times lower than the single-shot threshold value.

To summarize, we have optimized our earlier hybrid CARS experiment and demonstrated that the CARS spectrum from  $\approx 10^4$  bacterial endospores can be obtained in a single, femtosecond laser shot. This is to be compared with our initial acquisition time of many seconds. The dependence of the CARS signal on the energy of the ultrashort preparation pulses shows that it scales linearly, up to a peak energy fluence of  $\approx 0.2 \text{ J}/\text{cm}^2$  per pulse. Further signal increase is hampered by photodamage of the sample interacting with the strong electric field of the femtosecond pulses. Finally, it is important to note that spore detection, addressed in this article, is only one application of our technique. The technique can be adapted in a straightforward fashion for many other applications requiring chemical specificity and short acquisition times. A particularly interesting example is CARS microscopy (37–40).

## Methods

We used a Ti:Sapphire-based laser system (Legend; Coherent;  $\approx 805 \text{ nm}$ , 50 fs, 1 kHz rep rate, 1 mJ/pulse) pumping two optical parametric amplifiers (OPAs) (OPeRA; Coherent). The signal beams at two different wavelengths, generated in the OPAs by the down-conversion of 805-nm photons, provided pump ( $\lambda_1 \sim 1.25 \mu\text{m}$ ) and Stokes ( $\lambda_2 \sim 1.55 \mu\text{m}$ ) pulses. The probe pulses came from the residual 805-nm beam spectrally filtered and shaped by a 4f pulse shaper with a mechanical slit in its Fourier plane. The resulting pulse shape had a sinc-squared intensity profile. The three beams were arranged in folded-BOXCAR geometry and focused on the sample, as shown schematically in Fig. 1a. The FWHM diameters for the pump, Stokes, and probe beams on the sample were 36, 43, and 32  $\mu\text{m}$ , respectively (from knife-edge measurements). The timing between the three input pulses was adjusted by computer-controlled delay stages. The generated CARS signal was collected in the backward direction, at a  $30^\circ$  angle to the main axis, by a 2-inch-diameter lens (focal length of 10 cm) and directed into a spectrometer (Spectrograph-250i; Chromex) with an attached liquid nitrogen-cooled CCD (Spec-10; Princeton Instruments).

**ACKNOWLEDGMENTS.** This work was supported by Office of Naval Research Award N00014-03-1-0385, the Defense Advanced Research Projects Agency, National Science Foundation Grant PHY-0354897, an award from the Research Corporation, and Robert A. Welch Foundation Grants A-1261 and A-1547.

1. Pushkarsky MB, Webber ME, Macdonald T, Patel CKN (2006) High-sensitivity, high-selectivity detection of chemical warfare agents. *Appl Phys Lett* 88:044103.
2. Pushkarsky MB, et al. (2006) High-sensitivity detection of TNT. *Proc Natl Acad Sci USA* 103:19630–19634.
3. Edwards KA, Clancy HA, Baeumner AJ (2006) Bacillus anthracis: toxicology, epidemiology and current rapid-detection methods. *Anal Bioanal Chem* 384:73–84.
4. Spiro TG (1987) *Biological Applications of Raman Spectroscopy* (Wiley, New York).
5. Arp Z, Autrey D, Laane J, Overman SA, Thomas GJ (2001) Structural studies of viruses by Raman spectroscopy part LXXI—Tyrosine Raman signatures of the filamentous virus Ff are diagnostic of non-hydrogen-bonded phenoxyls: Demonstration by Raman and infrared spectroscopy of p-cresol vapor. *Biochemistry* 40:2522–2529.
6. Esposito AP, et al. (2003) Analysis of single bacterial spores by micro-Raman spectroscopy. *Appl Spectrosc* 57:868–871.
7. Chan JW, et al. (2004) Reagentless identification of single bacterial spores in aqueous solution by confocal laser tweezers Raman spectroscopy. *Anal Chem* 76:599–603.
8. Nelson WH, Dasari R, Feld M, Sperry JF (2004) Intensities of calcium dipicolinate and Bacillus subtilis spore Raman spectra excited with 244 nm light. *Appl Spectrosc* 58:1408–1412.
9. Farquharson S, et al. (2004) Detecting Bacillus cereus spores on a mail sorting system using Raman spectroscopy. *J Raman Spectrosc* 35:82–86.
10. Zhang XY, Young MA, Lyandres O, Van Duyne RP (2005) Rapid detection of an anthrax biomarker by surface-enhanced Raman spectroscopy. *J Am Chem Soc* 127:4484–4489.
11. Zhang XY, Zhao J, Whitney AV, Elam JW, Van Duyne RP (2006) Ultrastable substrates for surface-enhanced Raman spectroscopy: Al<sub>2</sub>O<sub>3</sub> overlayers fabricated by atomic layer deposition yield improved anthrax biomarker detection. *J Am Chem Soc* 128:10304–10309.
12. Maker PD, Terhune RW (1965) Study of optical effects due to an induced polarization third order in electric field strength. *Phys Rev* 137:A801.
13. Levenson MD, Bloembergen N (1974) Dispersion of nonlinear optical susceptibility tensor in centrosymmetric media. *Phys Rev B* 10:4447–4463.
14. Scully MO, et al. (2002) FAST CARS: Engineering a laser spectroscopic technique for rapid identification of bacterial spores. *Proc Natl Acad Sci USA* 99:10994–11001.
15. Hanna SF, et al. (2003) Electronic-resonance-enhanced coherent anti-Stokes Raman spectroscopy of nitric oxide. *Appl Phys Lett* 83:1887–1889.
16. Rebane A, et al. (2005) Femtosecond resonance enhanced CARS for background-free detection of organic molecules. *J Mod Opt* 52:1243–1253.
17. Rabitz H, de Vivie-Riedle R, Motzkus M, Kompa K (2000) Chemistry—Whither the future of controlling quantum phenomena? *Science* 288:824–828.
18. Voroshilov A, Otto C, Greve J (1997) On the coherent vibrational phase in polarization sensitive resonance CARS spectroscopy of copper tetraphenylporphyrin. *J Chem Phys* 106:2589–2598.
19. Muller M, Schins JM (2002) Imaging the thermodynamic state of lipid membranes with multiplex CARS microscopy. *J Phys Chem B* 106:3715–3723.
20. Chen JX, Volkmer A, Book LD, Xie XS (2002) Multiplex coherent anti-stokes Raman scattering microspectroscopy and study of lipid vesicles. *J Phys Chem B* 106:8493–8498.
21. Kee TW, Cicerone MT (2004) Simple approach to one-laser, broadband coherent anti-Stokes Raman scattering microscopy. *Opt Lett* 29:2701–2703.
22. Petrov GI, Yakovlev VV (2005) Enhancing red-shifted white-light continuum generation in optical fibers for applications in nonlinear Raman microscopy. *Opt Express* 13:1299–1306.
23. Materny A, et al. (2000) Wave packet dynamics in different electronic states investigated by femtosecond time-resolved four-wave-mixing spectroscopy. *Appl Phys B* 71:299–317.
24. Siebert T, Schmitt M, Grafe S, Engel V (2006) Ground state vibrational wave-packet and recovery dynamics studied by time-resolved CARS and pump-CARS spectroscopy. *J Raman Spectrosc* 37:397–403.
25. Lucht RP, Roy S, Meyer TR, Gord JR (2006) Femtosecond coherent anti-Stokes Raman scattering measurement of gas temperatures from frequency-spread dephasing of the Raman coherence. *Appl Phys Lett* 89:251112.
26. Pestov D, et al. (2007) Optimizing the laser-pulse configuration for coherent Raman spectroscopy. *Science* 316:265–268.
27. Prince BD, Chakraborty A, Prince BM, Stauffer HU (2006) Development of simultaneous frequency- and time-resolved coherent anti-Stokes Raman scattering for ultrafast detection of molecular Raman spectra. *J Chem Phys* 125:044502.
28. Kano H, Hamaguchi H (2006) Dispersion-compensated supercontinuum generation for ultrabroadband multiplex coherent anti-Stokes Raman scattering spectroscopy. *J Raman Spectrosc* 37:411–415.
29. Roy S, Meyer TR, Gord JR (2005) Time-resolved dynamics of resonant and nonresonant broadband picosecond coherent anti-Stokes Raman scattering signals. *Appl Phys Lett* 87:264103.
30. Evans CL, Potma EO, Xie XS (2004) Coherent anti-Stokes Raman scattering spectral interferometry: determination of the real and imaginary components of nonlinear susceptibility  $\chi^{(3)}$  for vibrational microscopy. *Opt Lett* 29:2923–2925.
31. Lim SH, Caster AG, Leone SR (2005) Single-pulse phase-control interferometric coherent anti-Stokes Raman scattering spectroscopy. *Phys Rev A* 72:041803.
32. Potma EO, Evans CL, Xie XS (2006) Heterodyne coherent anti-Stokes Raman scattering (CARS) imaging. *Opt Lett* 31:241–243.
33. Lim SH, Caster AG, Nicolet O, Leone SR (2006) Chemical imaging by single pulse interferometric coherent anti-Stokes Raman scattering microscopy. *J Phys Chem B* 110:5196–5204.
34. Oudar JL, Smith RW, Shen YR (1979) Polarization-sensitive coherent anti-Stokes Raman spectroscopy. *Appl Phys Lett* 34:758–760.
35. Oron D, Dudovich N, Yelin D, Silberberg Y (2002) Quantum control of coherent anti-Stokes Raman processes. *Phys Rev A* 65:043408.
36. Dudovich N, Oron D, Silberberg Y (2003) Single-pulse coherent anti-Stokes Raman spectroscopy in the fingerprint spectral region. *J Chem Phys* 118:9208–9215.
37. Duncan MD, Reintjes J, Manuccia TJ (1982) Scanning coherent anti-Stokes Raman microscope. *Opt Lett* 7:350–352.
38. Zumbusch A, Holtom GR, Xie XS (1999) Three-dimensional vibrational imaging by coherent anti-Stokes Raman scattering. *Phys Rev Lett* 82:4142.
39. Evans CL, et al. (2005) Chemical imaging of tissue in vivo with video-rate coherent anti-Stokes Raman scattering microscopy. *Proc Natl Acad Sci USA* 102:16807–16812.
40. Ogilvie JP, Beaurepaire E, Alexandrou A, Joffe M (2006) Fourier-transform coherent anti-Stokes Raman scattering microscopy. *Opt Lett* 31:480–482.